



Gummibers

Sample ID: 25814
Strain: Gummibers
Matrix: Plant
Type: Flower - Cured
Sample Size: ; Batch:

Produced:
Collected:
Received:
Completed: 11/05/2025
Batch#: 1

Client
Hemp Hills
Lic. #
NA
NA, CA 92705



Summary

Test	Date Tested	Result
Batch		Complete
Cannabinoids		Complete
Moisture		12.8% - Complete

Cannabinoids

Complete

13.838%

Total THC

1.856%

Total CBD

15.944%

Total Cannabinoids

Analyte	LOD	LOQ	Result	Result
	mg/g	mg/g	%	mg/g
CBC	0.009	0.025	0.1118	1.118
CBD	0.025	0.050	1.7977	17.977
CBDa	0.019	0.050	0.0664	0.664
CBG	0.025	0.050	0.1076	1.076
CBN	0.009	0.025	ND	ND
$\Delta 8$ -THC	0.019	0.050	ND	ND
$\Delta 9$ -THC	0.019	0.050	0.2540	2.540
THCa	0.013	0.025	15.4891	154.891
THCV	0.025	0.050	0.0303	0.303
Total THC			13.838	138.379
Total CBD			1.856	18.560
Total CBG			0.108	1.076
Total			15.944	159.436

Date Tested:

Total THC = THCa * 0.877 + $\Delta 9$ -THC + $\Delta 8$ -THC; Total CBD = CBDa * 0.877 + CBD; Total CBG = CBGa * 0.877 + CBG.

Total Cannabinoids = Total THC + Total CBD + Total CBG + minor cannabinoids.

Cannabinoids: HPLC, SOP-004

Water Activity: Water Activity Meter, SOP-012

Moisture Content: Moisture Analyzer, SOP-011

Foreign Matter: Visual Inspection, SOP-001

ND = Not Detected, NR = Not Reported, LOD = Limit of Detection, LOQ = Limit of Quantitation. This product has been tested by Excelbis Labs LLC using valid testing methodologies and a quality system as required by state law. All LQC samples were performed and met the prescribed acceptance criteria in 16 CCR section 5730, pursuant to 16 CCR section 5726(e)(13). Values reported relate only to the product tested. Excelbis Labs LLC makes no claims as to the efficacy, safety or other risks associated with any detected or non-detected levels of any compounds reported herein. This Certificate shall not be reproduced except in full, without the written approval of Excelbis Labs LLC. This Certificate of Analysis is limited to the sample tested in a batch. This Certificate does not make any representation or warranty for all Products within the tested Batch.